



Supervision directly by an individual knowledgeable in laser safety.

Class4 - Laser Safety Considerations
www.rayaphotonics.com

Require storage or disabling of lasers when not in use.

Entry of any noninvolved personnel requires approval.

All windows, doorways, open portals, etc., of an enclosed facility should be covered or restricted to reduce any escaping laser beams below appropriate ocular MPE level.

A beam stop of an appropriate material must be used to terminate all potentially hazardous beams.

The beam path of the laser must be located and secured above or below eye level for any standing or seated position in the facility.

Appropriate laser protective eye wear must be provided all personnel within the laser controlled area.

Use diffusely reflecting materials near the beam, where appropriate.

